

Material Description

Reference	CT21
Commercial name	General Electric Lightspeed 16
Characteristics	Lightspeed 16 slices from 2002 upgraded in Xstream console + Advantage workstation 4.2
Availability / Disponibilité	May 2012
Comments / Commentaires	SW version : 07MW11.10.H2_P_M16_G_Zeus Still used by customer. Covers will be repainted before expedition Under maintenance contract Tube Performix changed may 2012. Machine to be desinstalled 26th of may



System Information		
Item	Information	Status
Facility		
Site Name	LS16	-
System Type	LightSpeed16	-
Unique System Number		-
System ID		-
IP Address	172.17.147.30	Interface: eth0
Access Level	Non-proprietary	Valid
Software Installation Date	Thu 14 Apr 2011 09:15:19 AM	-
DASIS Camera	unknown	Not installed
DICOM Network Cameras	4 Camera(s)	Installed
Installed Tube	Performax	-
Tube Install Date	Mon Jul 25 15:44:31 2011	Total Patient Exams: 4487
Data Acquisition System	MDAS16	-
Power Distribution Unit	CT Compact	-

Current System Status		
Item	Information	Status
System Date	Mon, Feb 6, 2012	-
System Time	09:49:15 RET	-
Application Software	07MW11.10.H2.P.MI6.G.Zeus	running
Next Patient Exam	13297	-
SW Updates	Manual Installation Required	No

